

DBH
TLN

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 5/19/2015

Instrument Serial Number: 010562	As of 19-May-15
---	------------------------

Start Date	End Date	Agency Name
19-May-15		DFS Central Lab
16-Dec-14	19-May-15	Henrico County PD
04-Dec-14	16-Dec-14	DFS Central Lab
26-Jun-14	04-Dec-14	Culpeper PD
11-Jun-14	26-Jun-14	DFS Central Lab
05-Feb-13	11-Jun-14	New River Valley Regional Jail
26-Apr-12	05-Feb-13	DFS Central Lab
12-Mar-09	26-Apr-12	Norton PD
23-Jun-08	12-Mar-09	DFS Central Lab

DBH
TLN

INTOX EC/IR II **Quality Assurance Worksheet**

Instrument Serial Number 010562 Worksheet Start Date 5/19/2015

Location Henrico County Jail

Address 4301 E Parham Rd Richmond VA 23228

DFS Technician Donald Hall License No. 21250

☐ Laboratory ☒ On-Site

Site Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) 750 Reference Barometer (mm HG) 752

Reference Barometer(RB)Serial # 009109 RB Calibration Due 7/28/2015

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.298
0.300	0.296	0.287	0.304	Sample 2	0.297
Precision		sample min	sample max	Sample 3	0.297
0.001		0.297	0.298		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.098
0.100	0.099	0.096	0.102	Sample 2	0.098
Precision		sample min	sample max	Sample 3	0.098
0		0.098	0.098		


Dry gas standard Lot No. (with tank no.) AG316301-12

☐ Replaced dry gas standard (+O-ring)

☐ Installed at Location

☒ Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:



Notes:

5/19/2015

Instrument Serial Number

010562

Certification Date

☐ Calibrated☐ Certified☒ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

David B Hall

Date

5/19/15

Issuing Analyst

T. N. J.

Date

5/22/15

JPD




COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
--	--------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY HALL, DONALD, B		AGENCY DFS Central Lab
DFS LICENSE NUMBER 21250	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 05/19/2015
TEST EQUIPMENT NUMBER 010562		

RESULTS: TIME SAMPLE TAKEN 14:45 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____
SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____
OPERATOR'S SIGNATURE

DBH
TLN

IntoxNet MIS Report

Report Generated 19 May 2015 at 15:08

Test Results

Instrument Serial Number 010562

Test # 001148 Subject Test

Test Location 1 Department of
Test Date 19 May 2015

Test Location 2 Forensic Science
Test Time 14:38
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name HALL
Agency DFS Central Lab

Operator's First Name DONALD

Operator's Middle Initial B

Card Serial Number 121250

License Number 21250

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 19 May 2015 End Time 14:46

Result Time 14:45

Result Date 19 May 2015 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 14:38

Data Type BLK

Sample Value 0.000

Sample Time 14:39

Data Type CHK

Sample Value 0.099

Sample Time 14:40

Data Type BLK

Sample Value 0.000

Sample Time 14:41

Data Type SUBJ

Sample Value 0.000

Sample Time 14:42

Data Type BLK

Sample Value 0.000

Sample Time 14:43

Data Type SUBJ

Sample Value 0.000

Sample Time 14:45

Data Type BLK

Sample Value 0.000

Sample Time 14:46

Standard Type Dry Gas Std

Standard Value 0.099

Standard Lot Number AG316301-12

Standard Expiration Date 06/12/2015

Tank Pressure 154

Barometric Pressure 750 mmHg

Blow Sample Number 1 Blow Duration 3.47 sec

Blow Volume 1843 cc End-of-Blow Time 14:42

Blow Sample Number 2 Blow Duration 3.57 sec

Blow Volume 1735 cc End-of-Blow Time 14:45

Tamper Evident Stamp ad3dba4f

Test Status Code 0

Test Status Success

DBH
TLN

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010562 Test Number: 1149
Test Date: 05/19/2015 Test Time: 14:47 EDT
Dry Gas Target: 0.099
Lot Number: AG316301-12 Exp Date: 06/12/2015
Tank Pressure: 154 psi Barometric Pressure: 750 mmHg
System Check: Passed

Test	g/210L	Time
BLK	0.000	14:48
CHK	0.098	14:48
BLK	0.000	14:50
CHK	0.098	14:51
BLK	0.000	14:52
CHK	0.098	14:53

Test Status: Success

Calibration CRC: CE6A8082

WJ
TLN

Intox EC/IR-11: Accuracy Check

Department of Forensic Science

Serial Number: 010562 Test Number: 1150

Test Date: 05/19/2015 Test Time: 14:56 EDT

Dry Gas Target: 0.296

Lot Number: AG506501-10 Exp Date: 03/06/2017

Tank Pressure: 242 psi Barometric Pressure: 750 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	14:56
CHK	0.298	14:57
BLK	0.000	14:59
CHK	0.297	15:00
BLK	0.000	15:01
CHK	0.297	15:03

Test Status: *Success*

Calibration CRC: CE6A8082